

Manufacturing Quality

ISO 9001:2015

ISO 13485:2016

Certificate of Analysis

Onco-Ref™ Genomic DNA Reference Standard KIT c.2446G>T p.D816Y
 COSM1310


Product Information

Catalog Number:	ASE-6163G-1
Lot Number:	G163200723
Expiration Date:	2022-07-23
Quantity/Volume/Size:	1 µg / 20 µl / 1 vial
Gene Name:	KIT
Nucleotide Change:	c.2446G>T
Amino Acid Change:	p.D816Y
Mutation Id:	COSM1310
Cell:	HCT116
Storage Temperature:	-20°C
Shipping Temperature:	Ambient
Instructions for Use:	Refer to product Datasheet (Available www.accuref.com)
Intended Use:	Research Use Only. Not Intended for human or animal diagnostic or therapeutic use.

Quality Control

Test	Method	Acceptance Criteria	Result
Genotyping	Sanger Sequencing	Homozygous	PASS
Quality	Gel Electrophoresis	No Smear	PASS
Quantification	Qubit 4.0 Fluorometer	50 -60 ng /µl	PASS
DNA Purity	Nanodrop	A260/280 Ratio Between 1.8 to 2	PASS

AccuRef Diagnostics certifies that this standard meets the specifications stated in this certificate and warrants this product to meet the stated acceptance criteria through the expiration date when stored unopened as recommended.

	2020-10-15
Quality Assurance	Date

COA-ASE-6163G-1-V2