

Manufacturing Quality

ISO 9001:2015

ISO 13485:2016

Certificate of Analysis

**Onco-Ref™ Genomic DNA Reference Standard TP53 c.536A>G p.H179R
COSM10889**


Product Information

Catalog Number:	ASE-6198G-1
Lot Number:	G198191007
Expiration Date:	2021-10-06
Quantity/Volume/Size:	1 µg / 20 µl / 1 vial
Gene Name:	TP53
Nucleotide Change:	c.536A>G
Amino Acid Change:	p.H179R
Mutation Id:	COSM10889
Cell:	HCT116
Storage Temperature:	-20°C
Shipping Temperature:	Ambient
Instructions for Use:	Refer to product Datasheet (Available www.accuref.com)
Intended Use:	Research Use Only. Not Intended for human or animal diagnostic or therapeutic use.

Quality Control

Test	Method	Acceptance Criteria	Result
Quality	Gel Electrophoresis	No Smear	PASS
Quantification	Qubit 4.0 Fluorometer	50 -60 ng /µl	PASS
Genotype	NGS	Homozygous	PASS

AccuRef Diagnostics certifies that this standard meets the specifications stated in this certificate and warrants this product to meet the stated acceptance criteria through the expiration date when stored unopened as recommended.

	2019-10-17
Quality Assurance	Date

COA-ASE-6198G-1-V2