

Manufacturing Quality

ISO 9001:2015

ISO 13485:2016

## Certificate of Analysis

Onco-Ref™ Frozen Cell Pellet Reference Standard MET c.3743A>G p.Y1248C  
 COSM699


### Product Information

<b>Catalog Number:</b>	ASE-6300P
<b>Lot Number:</b>	E6300P291038
<b>Expiration Date:</b>	2019-09-26
<b>Quantity/Volume/Size:</b>	1,000,000 cells / 1 ml / 1 vial
<b>Gene Name:</b>	MET
<b>Nucleotide Change:</b>	c.3743A>G
<b>Amino Acid Change:</b>	p.Y1248C
<b>Mutation Id:</b>	COSM699
<b>Cell:</b>	HCT116
<b>Storage Temperature:</b>	-80°C
<b>Shipping Temperature:</b>	Dry Ice
<b>Instructions for Use:</b>	Refer to product Datasheet (Available <a href="http://www.accuref.com">www.accuref.com</a> )
<b>Intended Use:</b>	Research Use Only. Not Intended for human or animal diagnostic or therapeutic use.

### Quality Control

Test	Method	Acceptance Criteria	Result
Genotyping	Sanger Sequencing	Heterozygous	PASS

AccuRef Diagnostics certifies that this standard meets the specifications stated in this certificate and warrants this product to meet the stated acceptance criteria through the expiration date when stored unopened as recommended.

	<b>2018-10-03</b>
<b>Quality Assurance</b>	<b>Date</b>

COA-ASE-6300P-V2