

Manufacturing Quality

ISO 9001:2015

ISO 13485:2016

## Certificate of Analysis

Onco-Ref™ FFPE Cell Block Reference Standard RKO WT


### Product Information

<b>Catalog Number:</b>	ASO-4051
<b>Lot Number:</b>	O4051291919
<b>Expiration Date:</b>	2021-08-07
<b>Quantity/Volume/Size:</b>	24 mm (height) X 24 mm (length) X 4 mm (thickness); 7 mm diameter (cell core) / 1 block
<b>Cell:</b>	RKO
<b>Storage Temperature:</b>	-20°C
<b>Shipping Temperature:</b>	Ambient
<b>Instructions for Use:</b>	Refer to product Datasheet (Available <a href="http://www.accuref.com">www.accuref.com</a> )
<b>Intended Use:</b>	For Research Use Only. Not intended for human or animal diagnostics or therapeutics.

### Quality Control

Test	Method	Acceptance Criteria	Result
Quality (H&E)	H&E Staining	Stain Quality	PASS
Cell Density	Visual	>60%	PASS
Genotyping -WT	Whole Exon Sequencing	No New Mutation Detected	PASS

AccuRef Diagnostics certifies that this standard meets the specifications stated in this certificate and warrants this product to meet the stated acceptance criteria through the expiration date when stored unopened as recommended.

	<b>2018-10-17</b>
<b>Quality Assurance</b>	<b>Date</b>

COA-ASO-4051-V2